

This document is a glossary of symbols used in ManaMed packaging and labelling.

From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.1.1	Manufacturer	Indicates the medical device manufacturer
5.1.2 C EC REP	Authorized representative in the European Community/ European Union	Indicates the authorized representative in the European Community/ European Union
5.1.3	Date of manufacture	Indicates the date when the medical device was manufactured
5.1.4	Use-by date	Indicates the date after which the medical device is not to be used
5.1.5 LOT	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified
5.1.6 REF	Catalogue number	Indicates the manufacturer's catalogue number so that the medical device can be identified



From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.1.7	Serial number	Indicates the manufacturer's serial
Г		number so that a specific medical
		device can be identified
SN		
L J		
5.1.8	Importer	Indicates the entity importing the
		medical device into the locale
510	Distributor	Indicates the outity distribution the
5.1.9	Distributor	Indicates the entity distributing the medical device into the locale
		medical device into the locale
5.1.10	Model number	Indicates the <i>model number</i> or type
5.1.13 	Wodernamber	number of a product
		names of a product
11		
5.1.11	Country of	To identify the country of
¬	manufacture	manufacture of products
П		
5.3.1	Fragile, handle	Indicates a <i>medical device</i> that can
	with care	be broken or damaged if not
		handled carefully



From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.3.2	Keep away from	Indicates a medical device that
	sunlight	needs protection from light sources
5.3.3	Protect from heat and radio- active sources	Indicates a <i>medical device</i> that needs protection from heat and radioactive sources
5.3.4	Keep dry	Indicates a <i>medical device</i> that needs to be protected from moisture
_ J _		
5.3.5	Lower limit of temperature	Indicates the lower limit of temperature to which the <i>medical device</i> can be safely exposed
5.3.6	Upper limit of temperature	Indicates the upper limit of temperature to which the <i>medical device</i> can be safely exposed
5.3.7	Temperature limit	Indicates the temperature limits to which the <i>medical device</i> can be safely exposed



From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.3.8	Humidity limitation	Indicates the range of humidity to which the <i>medical device</i> can be safely exposed
5.3.9	Atmospheric pressure limitation	Indicates the range of atmospheric pressure to which the <i>medical</i> device can be safely ex- posed
5.4.1	Biological <i>risks</i>	Indicates that there are potential biological <i>risks</i> associated with the <i>medical device</i>
5.4.2	Do not re-use	Indicates a medical device that is intended for one single use only
5.4.3	Consult instructions for use or consult electronic instructions for use	Indicates the need for the user to consult the <i>instructions for use</i>
5.4.4	Caution	Indicates that caution is necessary when operating the device or control close to where the <i>symbol</i> is placed, or that the current situation needs operator awareness or operator action in order to avoid undesirable consequences



From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.4.5 LATEX	Contains or presence of natural rubber latex	Indicates the presence of dry natural rubber or natural rubber latex as a material of construction within the <i>medical device</i> or the packaging of a <i>medical device</i>
5.4.6	Contains human blood or plasma derivatives	Indicates a <i>medical device</i> that contains or incorporates human blood or plasma derivatives
5.4.8 BIO	Contains biological material of animal origin	Indicates a <i>medical device</i> that contains biological tissue, cells, or their derivatives, of animal origin
5.4.9 BIO	Contains biological material of human origin	Indicates a <i>medical device</i> that contains biological tissue, cells, or their derivatives, of human origin
5.4.10	Contains hazardous substances	Indicates a medical device that contains substances that can be carcinogenic, mutagenic, reprotoxic (CMR), or substances with endocrine disrupting properties
5.4.11	Contains nano materials	Indicates a medical device that contains nano materials



From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.4.12	Single patient multiple use	Indicates a medical device that may be used multiple times (multiple procedures) on a single patient
5.7.1	Patient number	Indicates a unique number associated with an individual patient
5.7.2	Patient name	Indicates the name of the patient
5.7.3	Patient identification	Indicates the identification data of the patient
5.7.4	Patient information website	Indicates a website where a patient can obtain additional information on the medical product
5.7.5 +	Health care centre or doctor	Indicates the address of the health care centre or doctor where medical information about the patient may be found



From ISO 15223-1 Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements		
Reference number and graphic	Title	Description
5.7.6	Date	Indicates the date that information was entered or a medical procedure took place
5.7.7	Medical device	Indicates the item is a medical device
5.7.8 「 A 文	Translation	Indicates that the original medical device information has undergone a translation which supplements or replaces the original information
5.7.9	Repackaging	Indicates that a modification to the original <i>medical device</i> packaging configuration has occurred
5.7.10	Unique device identifier	Indicates a carrier that contains unique device identifier information
5.2.7 NON STERILE	Non-sterile	Indicates a <i>medical device</i> that has not been subjected to a sterilization process



From ISO 7000 — Graphical symbols for use on equipment — Registered symbols Graphical symbols for use on equipment — Registered symbols		
Reference number and graphic	Title	Description
<u>3</u> 125	Normal hand washing,	To indicate that cleaning the
	maximum 40°C	textile article is allowed only by
1-()-1		washing by hand.
1 1 1 1 1 1 1 1 1 1		
\ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \		
3945	Mild hand washing, maximum	To indicate that only mild hand
	30°C	washing is allowed with the
\sim		maximum temperature of 30°C.
1 4111/ 7		
w		
3080	Elat devine	To indicate that flat drying is
3080 	Flat drying	
		allowed in the natural drying
		process.
 — 		
3103A	Line drying	To indicate that line drying is
Г	, 3	allowed in the natural drying
		process.
		p. 6 cc 5 5 .
3109	No tumble drying	To indicate that tumble drying is
		not allowed in the drying
		process.
	No. do planting	To be discussed in the control of th
3114	No dry cleaning	To indicate that dry cleaning is
		not allowed.
(\mathbf{X})		



From ISO 7000 — Graphical symbols for use on equipment — Registered symbols Graphical symbols for use on equipment — Registered symbols		
Reference number and graphic	Title	Description
3124	No bleaching	To indicate that bleaching the textile article is not allowed.
3113	No ironing	To indicate that ironing is not allowed.
M002	Refer to instruction manual/booklet	To signify that the instruction manual/booklet must be read
0434A	Caution	To indicate that caution is necessary when operating the device or control close to where the symbol is placed. To indicate that the current situation needs operator awareness or operator action in order to avoid undesirable consequences. On the health app quality label: to indicate that the health app requires approval from a health professional for use.
5172	Class II equipment	To identify equipment meeting the safety requirements specified for Class II equipment according to IEC 61140.
5840	Type B applied part	To identify a type B applied part complying with ,IEC 60601-1.



From ISO 7000 — Graphical symbols for use on equipment — Registered symbols Graphical symbols for use on equipment — Registered symbols		
Reference number and graphic	Title	Description
5333	Type BF applied part	To identify a type BF applied part complying with IEC 60601-1.

Other Symbols		
Graphic	Description	Reference
CE	European Conformity (CE) Mark	European Medical Device Directive 93/42/EEC
LATEX	Indicates that the product is latex-free	N/A
Ronly	Prescription only	N/A